

Municipiul Lugoj
Serviciul Asistenta Medicala Invatamant
Nr.inreg.119/19.05.2020

Aprobat,
Primar,
Prof.Ing.Francisc Boldea

Caiet de sarcini

1.TERMOMETRE NON CONTACT

- termometre digitale cu infrarosu non contact
- distanța de termometrizare 5-10cm.
- eroare posibilă-0,1-0,2
- nr.termometre-14buc.

Medic coordonator,
Dr.Stan Didina

